



# **The Notified Body Opinion**

An assessor's view

Maeve Lally, EMA Drug Device Guideline Drafting group, BWP Member

Webinar on the implementation of Article 117 of the Medical Device Regulation

27 November 2020





#### Disclaimer

The contents of this presentation are personal observations and are not necessarily representative of the HPRA, EMA or any other agency





#### Regulation 2017/745

#### Article 117

#### Amendment to Directive 2001/83/EC

In Annex I to Directive 2001/83/EC, point 12 of Section 3.2. is replaced by the following:

'(12) Where, in accordance with the second subparagraph of Article 1(8) or the second subparagraph of Article 1(9) of Regulation (EU) 2017/745 of the European Parliament and of the Council (\*), a product is governed by this Directive, the marketing authorisation dossier shall include, where available, the results of the assessment of the conformity of the device part with the relevant general safety and performance requirements set out in Annex I to that Regulation contained in the manufacturer's EU declaration of conformity or the relevant certificate issued by a notified body allowing the manufacturer to affix a CE marking to the medical device.

If the dossier does not include the results of the conformity assessment referred to in the first subparagraph and where for the conformity assessment of the device, if used separately, the involvement of a notified body is required in accordance with Regulation (EU) 2017/745, the authority shall require the applicant to provide an opinion on the conformity of the device part with the relevant general safety and performance requirements set out in Annex I to that Regulation issued by a notified body designated in accordance with that Regulation for the type of device in question.

Applicant shall provide an opinion on the conformity of the device part with the relevant GSPRs as set out in Annex I - NBOp





# **Medicines Authority - Assessor Remit**

- In accordance with Directive 2001/83/EC assessment of any risk relating to quality, safety or efficacy of the medicinal product
- As per draft guideline (EMA/CHMP/QWP/BWP/259165/2019):

"The assessment of the suitability of a device for its intended purpose should take into account the relevant quality aspects of the device itself and the context of its use with the medicinal product"

"Complexity of the device, patient characteristics, user requirements, clinical setting/ use environment"

- Aim to
- a) Ensure the B/R (Q/S/E) of the finished medicinal product is positive
- b) Minimise duplication of effort for regulatory authority/ notified body/ applicants



# What we suggested:



#### 734 Annex 2: Template cover sheet for Notified Body Opinion

It is intended that this document is completed to provide information on the medicinal product containing an integral medical device. Where an application is made for a DDC containing an individual device part, the applicant or MAH completes this section.

#### GENERAL INFORMATION

PRODUCT DETAILS		
Invented / Trade name of the medicinal product	<as maa="" per=""></as>	
Applicant	<name address="" and="" entity<br="" i.e.="" legal="" mah="" of="">holding the MA&gt;</name>	
Marketing authorisation type	<e.g. application="" centralised=""></e.g.>	
Marketing authorisation procedure number	<g.g. eu="" xxx="" xxx.=""></g.g.>	
Pharmaco-therapeutic group (ATC code)	<e.g. c52="" d08a=""></e.g.>	
Indication(s)	<as per="" spc4.1=""></as>	
Pharmaceutical form(s) and strength(s)	<e.g. (collection="" 10mg,="" 20mg="" for="" injection,="" pre-filled="" pring=""></e.g.>	
INN (or common name) of the active substance(s):	<astrer marx<="" td=""></astrer>	
Authorisation to use NROp	Jitably authorised / signed by the MAH*	

Official address Domenico Sociationes 6 • 1083 HS Amsterdam • The Netherlands
Address for visits and deliveries. Refer to www.ama.europa.au/how-to-find-us
Send us a question (5 to towww.ama.europa.eu/contact Telephone +31 (0)85 781 6000

An apincy of the Europa

© European Medicines Agency, 2020. Reproduction is authorised provided the source is acknowledged





	Date Version Number
	NB logo
	NB name and address
	NB number
	. •
	Notified Body Opinion
	(Article 117 of the Medical Device Regulation (EU 2017/745)
	Compliance of device(s) incorporated into an integral drug-device combination product
	with.
	Annex I (General Safety and Perform nor Requirements)
	Medical Device Regulation (5U.2) (6)
	-MP
	Administrative reference number:
	10 50
	Administrative refer and number:
	-04
1	
	Authorised by (name):
	Authorised by (signature):
	Issue date (YYYY/MM/DD):
	* This is a suggested template, to facilitate ease of review by the Competent Authority; others can be
	used

/OWP/BWP/259165/2019





# **Product 1 – Medicinal product dossier overview (eCTD)**

Product 1 – Biological (NAS) in PFS, combined with either safety syringe (SS) or auto-injector (AI)

#### Location of device information within the dossier

- 3.2.P.1 brief description
- 3.2.P.2 design verification of devices as per ISO, design validation including summary of HF studies, clinical studies
- 3.2.P.3 assembly process and controls
- 3.2.P.5 device-specific specifications
- 3.2.P.7 diagrams, material of construction, critical material attributes
- 3.2.P.8 long-term and accelerated study results and protocols
- 3.2.R Human Factors and clinical study summary, NBOp for SS and for Al





### **Product 1 - NBOp**

- Two NBOps (SS and AI) each < 30 pages</li>
- Did not use EMA draft template
- Clearly stated MAH, product, internal reference number
- Overall conclusion clearly stated met requirements of Article 117,
   reviewed technical documentation against Annex I of Reg (EU) 2017/745
- Includes general description aligned with module 3, intended purpose of DDC aligned with SPC, assessment of GSPRs (tabular format, clearly highlights yes/no/partial/na and brief summary of the data reviewed and overall conclusion

or cells, or their derivatives, of animal origin   □PARTIAL  410/01 rev 03 have been met
--





#### **Product 2 – Medicinal product dossier overview (eCTD)**

- MAb (biosimilar) in PFS, assembled with SS or Al Location of device information within the dossier
- 3.2.P.1 brief description
- 3.2.P.2 design verification plan and verification report, scale up, transport
- 3.2.P.3 assembly process and controls, definition of batch size
- 3.2.P.5 device-specific specifications
- 3.2.P.7 diagrams, material of construction, specifications
- 3.2.P.8 based on primary container closure
- 3.2.R NBOp for SS and for AI, device RMP, URRA, threshold analysis
- Module 2.5 and 2.7 HF and Usability studies
- 1 x NBOp under MDR, 1 x ER checklist and design file under MDD not required at present





### **Product 2 - NBOp**

- ~ 30 pages
- Did not use EMA draft template
- Clearly stated MAH, product, internal reference number
- Introduction clearly identified GSPRs which could not be confirmed gaps were identified and highlighted to CA where information had not been provided, useful for targeted assessment
- Body of report product description, purpose of DDC, therapeutic context, list of GSPRs in tabular format

		I	<u>-</u>
7	Design & manufacture – transport, storage	□YES ⊠NO □PARTIAL	Transit studies have not been completed and therefore documentation was not available to demonstrate the suitability of the packaging for transport.
			I





# **Product 3 - Medicinal product dossier overview (eCTD)**

- MAb (line extension) in PFS assembled with NSP device or Al Location of device information within the dossier
- 3.2.P.1 brief description
- 3.2.P.2 RMP, design verification plan report, design validation and HF program, bridging clinical to commercial
- 3.2.P.3 assembly process and controls
- 3.2.P.5 device-specific specifications,
- 3.2.P.7 technical drawings, material of construction, specifications
- 3.2.P.8 platform approach to stability
- 3.2.R certificates of conformance, GSPRs assessment by MAH for NSP-PFS & AI, NBOp for PFS
- Module 5.3 Other studies HF engineering and HF summary reports





### **Product 3 - NBOp**

- Two NBOps (80+, 100+ pages)
- Did not use EMA draft template
- Clearly stated MAH, product, EU number, internal ref number
- Overall conclusion clearly stated

The device conforms to the relevant <b>General Requirements</b> as outlined in <b>Chapter I</b> of Annex I of Regulation (EU) 2017/745	⊠ Yes	□No	☐ Partially
The device conforms to the relevant Requirements regarding Design and Manufacture as outlined in Chapter II of Annex I of Regulation (EU) 2017/745	⊠ Yes	□No	☐ Partially
The device conforms to the relevant Requirements regarding the Information supplied with the Device as outlined in Chapter III of Annex I of Regulation (EU) 2017/745	⊠ Yes	□No	☐ Partially

• Includes general description aligned with module 3, intended purpose of DDC aligned with SPC, classification of device, assessment of GSPRs (tabular format, clearly highlights yes/no/partial and brief summary of the data reviewed and overall conclusion



# Summary of NBOp differences



	NBOp 1	NBOp 2	NBOp 3	NBOp 4
Length	< 30 pages	30+ pages	80+ pages	120 +pages
MAH, product name, reference number	V	V	V	V
Description aligned with M3	√	X	V	V
Intended purpose aligned with SPC	V	X	V	V
Device classification	X	X	V	V
Alignment with GSPRs		X	V	V
Clear opinion	$\checkmark$	$\checkmark$	V	V
List of standards	$\checkmark$	$\checkmark$	X	X
Summary of risk assessment	X	X	V	V





### Key take home messages

#### For MAH

- Include relevant information in dossier, consistent with draft guideline
- Clearly cross reference to other sections of module 3/ other sections of dossier for ease of review e.g. location of HF study

#### For NB in NBOp

- Clear conclusions at beginning, highlight areas which were not assessed/ not met which should be focused on during MAA
- Consistency between NBOp and dossier/ SPC
- GSPRs in tabular format with explanation for decision
- Medical device classification is useful





#### **Conclusion**

- If MAH has clearly presented device data in the dossier
   AND
- Conclusions of NBOp are clearly outlined
   THEN
- Very little overlap/duplication of effort
- Assessors cross-check device details match SPC/ dossier,
- Assess information relevant to device in combination with the medicinal product
- Focus on gaps identified/ non-conformance with GSPRs