

1 December 2025 Rev.6
EMA/37991/2019
Human Medicines Division

Questions & Answers for applicants, marketing authorisation holders of medicinal products and notified bodies regarding medicines used in combination with medical devices and consultation procedures for certain medical devices

This Question and Answer (Q&A) document provides practical considerations concerning the application of the medical devices and the *in vitro* diagnostic medical devices regulations in the context of combinations of medicinal products with medical devices.

This document has been produced to provide guidance to applicants, marketing authorisation holders (MAH) and notified bodies (NB) as regards aspects falling within the scope of the European Medicines Agency's activities and should be read in conjunction with the medical devices [Regulation \(EU\) 2017/745 \(MDR\)](#), the *in vitro* diagnostic medical devices [Regulation \(EU\) 2017/746 \(IVDR\)](#) and the Medical Device Coordination Group ([MDCG](#))¹ [guidance documents](#). The MDR and IVDR replace the three Directives (90/385/EEC, 93/42/EEC and 98/79/EC) for medical devices. The MDR came into application on 26 May 2021 but provides for a transitional period for certain devices. The IVDR came into application on 26 May 2022 but also provides for a transitional period for certain devices. These regulations include provisions concerning the responsibilities of the EMA, National Competent Authorities (NCA) for medicinal products and medical devices and notified bodies as regards combinations of medicinal products with medical devices as follows:

- For medical devices that form an integral product with a medicinal product (Regulation (EU) 2017/745, second subparagraph of Article 1(8) and 1(9)), new requirements to provide an EU declaration of conformity issued by the medical device manufacturer, or a certificate of conformity (so-called hereafter EU certificate) or an opinion from a notified body designated under Regulation (EU) 2017/745² for the type of device in question are applicable in certain circumstances (Art. 117).
- For medical devices incorporating a medicinal substance with action ancillary to the device, Regulation (EU) 2017/745 Article 1 (8), the notified body must seek a scientific opinion from either an NCA or EMA³. The notified body must seek the opinion of EMA for medicinal products falling

¹ The [MDCG](#) is the Member State group responsible for the oversight of implementation of the medical device Regulations. It is set up according to Art. 103 of Regulation (EU) 2017/745 and Art. 98 of Regulation (EU) 2017/746.

² Regulation 2017/745 Article 117

³ Regulation 2017/745 Annex IX 5.2

exclusively within the scope of centralised procedure⁴, or that incorporate human blood or plasma derivatives.

- For medical devices that are composed of substances, or of combinations of substances, that are systemically absorbed by the body in order to achieve their intended purpose, the notified body must seek a scientific opinion from either an NCA or EMA (Regulation (EU) 2017/745 Article 52(11)⁵.
- For companion diagnostics, the notified body must seek a scientific opinion from either an NCA or EMA (Regulation (EU) 2017/746, Article 48(3) and (4)⁶).

This document covers guidance for:

- Medical devices that form an integral product with a medicinal product,
- Medicinal products that include a medical device in the secondary packaging of the medicinal product ('co-packaged'),
- Consultation procedure for ancillary medicinal substances that are integral part of medical devices,
- Consultation procedure for companion diagnostics.

This "questions and answers" document provides regulatory and procedural guidance and should be read in conjunction with the EMA [Guideline on quality documentation for medicinal products when used with a medical device](#). This guideline describes the information that should be presented in the Quality part of a marketing authorisation dossier for a medicinal product when it is used with a medical device, or device part, and submitted in accordance with Directive 2001/83/EC and/or Regulation (EC) No 726/2004. This guideline focuses on product-specific quality aspects of a medical device, or device part, that may have an impact on the quality, safety and/or efficacy (and hence overall benefit/risk determination) of a medicinal product.

This "questions and answers" document is being updated continuously and will be marked by "New" or "Rev." with the relevant date upon publication.

⁴ Annex I, Regulation (EC) No 726/2004

⁵ Regulation 2017/745 Annex IX 5.4

⁶ Regulation 2017/746 Annex IX 5.2, Annex X 3(k)

Table of contents

1. Combinations of medicinal products and medical devices	5
1.1. What regulatory framework does a product incorporating both medicinal product and medical device fall under? Rev. May 2024	5
1.2. How do I choose a notified body for my integral drug-device combination? Rev. May 2024	6
1.3 How to obtain advice on the qualification/classification of my drug-device combination, especially for borderline products? New May 2024.....	6
2. Integral Drug-Device Combinations.....	7
2.1. When is my medical device considered to form an integral product with a medicinal product? Rev. May 2024	7
2.2. What is Article 117 and what does it mean for medicinal products? Rev. January 2025 .	8
2.3. At what stage of the MAA do I need to submit the notified body opinion? Rev. May 2024	9
2.4. Can I provide a notified body opinion or MAH's GSPRs compliance statement for class I concluding on partial compliance with the GSPRs? What is the scope of the notified body opinion? New January 2025.....	10
2.5. How does Article 117 of the MDR impact iDDCs authorised under MDD? Rev. December 2025	10
2.6. Is it possible to submit a notified body certificate issued under the Directives (90/385/EEC or 93/42/EEC) to comply with Article 117? Rev. May 2024	10
2.7. How will the notified body opinion or MAH's GSPRs compliance statement be reflected in the European Public Assessment Report (EPAR)? Rev. January 2025	10
2.8. What is the impact of the MDR and Article 117 on marketing authorisation applications of an iDDC on a Mutual Recognition Procedure submitted on or after the 26 May 2021? Rev. January 2025	11
2.9. Do the requirements of MDR Article 117 also apply to an application for medicinal products to be used outside of the European Union (Article 58 or EU-M4all)? Rev. June 2021	11
2.10. Are the requirements for UDI (unique device identifier) applicable to a medicinal product that incorporates, as an integral part, a medical device? Rev. May 2024	12
3. Medicinal products that include a medical device in the secondary packaging of the marketed medicinal product (co-packaged)	12
3.1. How will the MDR affect the co-packaged medical device? Rev. May 2024	12
3.2. What requirements for medical device labelling are applicable to medical devices "co-packaged" with medicinal products? Rev. May 2024	12
3.2.1 If co-packaged medical devices class I and class IIa, are supplied without an individual packaging and it is not technically feasible to implement the labelling requirements on the device itself, what alternative solutions could be considered to display the labelling requirements? New May 2024	13
3.3. Do I need to submit a declaration of conformity / EU certificate as part of the dossier for a co-packaged medical device? Rev. May 2024	14
3.4. What actions, if any, do I need to take if my co-packaged device is up-classified and requires to be certified by a notified body for the first time? Rev. May 2024	15
4. Lifecycle management.....	15
4.1. Will I need to provide a (new or updated) EU declaration of conformity/certificate of conformity issued by a notified body/notified body opinion if there are changes to the device (or device part) after the initial marketing authorisation? Rev. December 2025	15

4.2. How should I submit changes to the terms of the Marketing Authorisation following changes to the device (or device part)? Rev. December 2025 16

5. Consultation procedure for ancillary medicinal substances in medical devices (MDR Art 1(8)) 19

5.1. What type of consultation procedure needs to be carried out for an ancillary medicinal substance that has already been consulted under the medical device Directive 93/42/EEC? Rev. May 2024 19

6. Consultation procedure for companion diagnostics 20

6.1. What type of consultation procedure needs to be carried out for a companion diagnostic New May 2024..... 20

1. Combinations of medicinal products and medical devices

1.1. What regulatory framework does a product incorporating both medicinal product and medical device fall under? *Rev. May 2024*

Products which combine a medicinal product (or substance) and a medical device are regulated either by Regulation (EU) 2017/745 on medical devices or Directive 2001/83/EC relating to medicinal products for human use, depending on their principal mode of action. Certain combinations of medicinal products and medical devices are governed by or require consultation of EMA or a national competent authority for medicinal products (NCA), as laid down in Articles 1(8) and 1(9) of the MDR.

The regulatory framework for devices incorporating medicinal substances as an integral part is laid down in Article 1(8) of MDR:

- Where the action of the medicinal substance is **ancillary** to the action of the medical device, the product is regulated as a medical device and must be CE marked. For these combinations, a scientific opinion on the quality and safety of the ancillary substance including the benefit or risk of its incorporation in the device must be provided from one of the national competent authorities for medicinal products or from the EMA (referred to collectively as medicines authority) before a notified body can issue a EU certificate. For more information on the consultation procedure by EMA and for a list of products previously reviewed by EMA, please see EMA webpage on [Ancillary medicinal substances in medical devices](#).
- Where the action of the medicinal substance is **principal** and not ancillary to the action of the medical device, the integral drug device combination (iDDC) is governed under the medicinal products framework and are referred to as iDDCs in the Q&A hereafter⁷. In that case, the relevant general safety and performance requirements of the Annex I of the MDR apply to the device part of the iDDC.

The regulatory framework for medical devices intended to administer medicinal products is laid down in Article 1(9) MDR:

- If the medicinal product and administration device are marketed as a **single integral product** intended **exclusively for use in the given combination** and is **not reusable**, the iDDC is governed under the medicinal products framework. In that case, the relevant general safety and performance requirements of Annex I of the MDR apply to the device part.
- In all other cases (e.g. where the medical device is co-packaged with the medicinal product or when the product information of the medicinal product refers to a specific device to be used and the device is obtained separately), the administration device is governed by the medical device framework. These administration devices must meet the requirements of the MDR and will need to be CE marked.

For the purpose of this document, integral products falling within MDR Article 1(8) second subparagraph and 1(9) second subparagraph, and for which the principal mode of action is pharmacological, metabolic or immunological, are regulated under the medicinal products framework.

There are cases where a medicinal product and a medical device are placed on the market in the same secondary packaging but do not form an integral product, for example, a vial containing a medicinal

⁷ so-called 'integral medicinal products' in the [EMA Guideline on quality documentation for medicinal products when used with a medical device](#) (EMA/CHMP/QWP/BWP/259165/2019)

product solution with an (empty) CE marked sterile syringe. These are referred as a medicine with a co-packaged device. This product is not considered an iDCC as the medical device falls under the first subparagraph of Article 1(9) of the MDR. Furthermore, when a medical device is referenced in the product information of a medicinal product, for a use in combination with the medicinal product, but the two products are not placed on the EU market within the same secondary packaging, requirements for these products are covered in section 3 of this Q&A.

Of note, IVD kits may not include medicinal products. If an IVD or an IVD kit is intended to be used with a medicinal product and the products are co-packaged, the combination with the medicinal product may not be qualified as an IVD kit. Each product must comply with its corresponding legislation.

1.2. How do I choose a notified body for my integral drug-device combination? *Rev. May 2024*

A notified body within the European Union (EU) is a third-party conformity assessment body designated, by an EU Member State authority, to carry out conformity assessment activities regarding medical devices and/or *in-vitro* diagnostics (IVDs) in accordance with the MDR and/or IVDR before being placed on the Union market. Companies are free to choose any notified body; the only criterion is that the notified body must be designated to carry out the conformity assessment procedure for the particular medical device for which a certification or notified body opinion is sought; this is defined by the notified body's designation codes. Applicants can check the [NANDO](#) (New Approach Notified and Designated Organisations) website, by clicking on 'Legislation' and select the relevant Directive/Regulation to search for a designated notified body according to the codes/scope needed.

1.3 How to obtain advice on the qualification/classification of my drug-device combination, especially for borderline products? *New May 2024*

Borderline products are those where it is not clear which regulatory framework applies, i.e., whether they fall under the medical device framework or the pharmaceutical framework.

The demarcation between the MDR on one hand and the pharmaceutical legislation on the other hand is crucial for applying the appropriate set of requirements.

The qualification/classification of a borderline product (e.g. whether it should be regulated as a medicinal product or as a medical device) lies with the national competent authorities for medicinal products and/or medical devices.

Several provisions to establish the demarcation between the two legal frameworks have been laid down in the MDR and the Directive 2001/83/EC and explanations and examples clarifying these provisions can be found in [MDCG 2022 – 5 Guidance on borderline between medical devices and medicinal products under Regulation \(EU\) 2017/745 on medical devices](#). In addition, the MDCG [Manual on borderline and classification for medical devices under Regulation \(EU\) 2017/745 on medical devices and Regulation \(EU\) 2017/746 on *in vitro* diagnostic medical devices](#) records the views of the Member State members of the MDCG Borderline and Classification Working Group regarding the regulatory status of a product (see Section 1.1.2 regarding the borderline between medical devices and medicinal products).

Qualification and risk classification of medical devices should be agreed with the notified body, if applicable. Any dispute between the manufacturer and the notified body regarding the classification of

a device is to be referred for a decision to the national competent authority for medical devices in the Member States in which the manufacturer has its registered place of business (Article 51(2) MDR and 47(2) IVDR).

In order to facilitate the development of innovative drug-device combinations and foster an early dialogue, informal input can be requested from the Innovative Task Force (ITF) of the European Medicines Agency. While the ITF may express scientific views on borderline products classification, it is the responsibility of the Applicant/MAH to contact a national competent authority for possible formal advice on the qualification of combination or borderline products and/or risk classification of a medical device, when needed and before submitting an application.

References

- [EU National Competent Authorities for human medicinal products](#)
- [EU National Competent Authorities for medical devices](#)
- [EMA guidance on how to request an ITF meeting](#)
- [MDCG borderline guidance and manual](#)

2. Integral Drug-Device Combinations

2.1. When is my medical device considered to form an integral product with a medicinal product? Rev. May 2024

There are two types of iDDCs⁸ according to MDR Articles 1(8) and 1(9) (see also Question 1.1).

1. A medical device that incorporates, as an integral part, a substance which, if used separately, would be considered a medicinal product and where the action of that substance is principal, the integral product will be regulated as a medicinal product (second subparagraph of Article 1(8)). Examples include an ingestible sensor that is incorporated into a medicinal product.
2. If a medical device used to administer a medicinal product is placed on the market in such a way that the two constituents parts (the medical device and medicinal product) form a single integral product which is intended exclusively for use in the given combination and which is not reusable, that single integral product shall be governed by Directive 2001/83/EC or Regulation (EC) No 726/2004, as applicable. The second paragraph of Article 1(9) of the MDR sets out three cumulative conditions that need to be satisfied at the moment of placing on the market:
 - i. the device and the medicinal product form a single integral product when placed on the market,
 - ii. the single integral product is intended exclusively for use in the given combination,
 - iii. the single integral product is not reusable.

Below are provided some examples.

For medicinal products meeting either one or both of the above definitions, the integral product shall be governed by Directive 2001/83/EC or Regulation (EC) No 726/2004, as applicable. However, the

⁸ so-called 'integral medicinal products' in the [EMA Guideline on quality documentation for medicinal products when used with a medical device](#) (EMA/CHMP/QWP/BWP/259165/2019)

relevant general safety and performance requirements set out in Annex I to Regulation (EU) 2017/745 shall apply as far as the safety and performance of the device part of the integral product is concerned.

Examples of integral products which are not reusable are pre-filled syringes, pre-filled pens, nebulisers pre-charged with a specific medicinal product, nasal and oromucosal sprays, patches for transdermal drug delivery, pre-filled inhalers and co-packaged pre-filled syringe with solvent (e.g. water for injections) used for administration of the reconstituted medicinal product.

In certain cases, the following components would be assessed by competent authorities in accordance with the requirements of:

- a container closure system e.g. eye drops nozzle without measuring function, syringe for reconstitution (without purpose for administration of the medicinal product) or syringe without measuring function specifically intended to transfer into an IV bag or
- excipients used in the manufacture of the finished medicinal product e.g. in some implants and some transdermal patches (using passive diffusion),

Note: the above examples are provided as a guide, however some devices could be classified differently due to additional features and functionality. If in doubt over the qualification of your product as a medicinal product, device or iDDC, it is recommended that you consult a national competent authority.

References

- [National Competent Authorities for human medicinal products](#)
- [National Competent Authorities for medical devices](#)
- [MDCG borderline guidance and manual](#)

2.2. What is Article 117 and what does it mean for medicinal products? *Rev. January 2025*

Applications for a marketing authorisation of an iDDC submitted as of 26 May 2021 must comply with the requirements of Annex I to Directive 2001/83/EC, point 12 of section 3.2, as amended by Article 117 of Regulation (EU) 2017/745.

Article 117 of Regulation (EU) 2017/745 sets out that the DDC marketing authorisation dossier must include, where available, the results of the assessment of conformity for the device part (i.e. the declaration of conformity or the relevant EU certificate issued by a notified body).

If the dossier does not include the results of the assessment of conformity, and an EU certificate from a notified body would be required if the device was used separately, then the applicant will be required to provide an opinion from a notified body on the conformity of the device part with relevant general safety and performance requirements (GSPRs) of Annex I to Regulation (EU) 2017/745 as part of the Marketing Authorisation Application. When a Declaration of Conformity is not available for class I (excluding Is and Im) medical devices, a MAH's statement of compliance with the relevant GSPRs of the MDR Annex I (to be listed) can be acceptable.

Article 117 applies to iDDCs referred to under the second subparagraphs of MDR Articles 1(8) and 1(9).

Article 117 does not apply in the case of combined advanced therapy medicinal products as defined under Article 2(1)(d) of Regulation (EC) No 1394/2007.

The submission of evidence of compliance with the general safety and performance requirements of the MDR in the marketing authorisation dossiers of integral DDCs did not apply to MAAs submitted before 26 May 2021, i.e. the date of application of the MDR.

In summary, marketing authorisation applications for an iDDC submitted as of 26 May 2021, must demonstrate that the device part meets the relevant requirements of Annex I of Regulation (EU) 2017/745 as presented in the below table.

Additional information may be requested to support the review of the benefit/risk assessment of the medicinal product in order to ensure a safe and effective use of the iDDC.

Table 1. Summary of changes for Marketing Authorisation Applications involving iDDCs

Risk class of medical device	New submissions as of 26 th May 2021
Class I sterile, measuring or reusable surgical instrument* Class IIa Class IIb Class III	The marketing authorisation dossier must include an EU certificate issued by a notified body designated for the type of device part in question. If the abovementioned documentation is not available, then an opinion** from a notified body must be provided for the medical device.
Class I non-sterile, non-measuring, or non-reusable surgical instrument	The marketing authorisation dossier must include a Declaration of Conformity for the medical device, or as an alternative for class I (excluding Is and Im) medical devices, when a Declaration of Conformity is not available, the MAH may submit a MAH's statement of compliance with the relevant GSPRs of the MDR Annex I (to be listed).
* the reader should note that iDDC as referred to in second subparagraph of Regulation 2017/745 Article 1(9) are not reusable **opinion on the conformity of the device part with the relevant general safety and performance requirements (GSPRs) set out in Annex I to Regulation 2017/745	

2.3. At what stage of the MAA do I need to submit the notified body opinion? *Rev. May 2024*

EMA/NCAs strongly recommend submitting the declaration of conformity / EU certificate / notified body opinion at the time of submission of the dossier of the initial marketing authorisation application for the medicinal product to facilitate a smooth running of the procedure. In case the applicant cannot provide the required documentation at the time of MAA submission, the relevant documents must be provided before an opinion on the medicinal product application can be issued. Applicants should discuss their plans to provide the required documentation during the EMA/NCA pre-submission meeting. The absence of the required documentation may result in additional clock stops during the procedure.

2.4. Can I provide a notified body opinion or MAH's GSPRs compliance statement for class I concluding on partial compliance with the GSPRs? What is the scope of the notified body opinion? *New January 2025*

A notified body opinion or for class I, a MAH's GSPRs compliance statement concluding on partial compliance cannot be accepted as the CHMP and the NCAs do not have the remit of assessing compliance of the device part with the General Safety and Performance requirements (GSPRs). This can hence lead to an evaluation issue and an updated notified body opinion confirming full compliance with the relevant GSPRs will need to be provided prior to the issuance of an opinion on the medicinal product.

Furthermore, the scope of the notified body opinion should correspond to the intended purpose of the device for the particular marketing authorisation application.

2.5. How does Article 117 of the MDR impact iDDCs authorised under MDD? *Rev. December 2025*

Annex I to Directive 2001/83/EC, point 12 of section 3.2, as amended by Article 117 of MDR, is not intended to apply retrospectively to iDDCs already authorised or to those MAAs that have been already submitted prior to 26 May 2021.

However, if after the granting of the marketing authorisation there is a major change to a device such as in its design, performance, safety or intended purpose of the device (part) that may have a significant impact on the delivery or the quality, safety, or efficacy of the medicinal product, or a new device is introduced, any required declaration of conformity / EU certificate / notified body opinion (or MAH's GSPRs compliance statement for class I) should be submitted as part of the appropriate regulatory procedure to EMA/NCA (see also Q4.1 and Q4.2).

2.6. Is it possible to submit a notified body certificate issued under the Directives (90/385/EEC or 93/42/EEC) to comply with Article 117? *Rev. May 2024*

A medical device with a certificate that was issued in accordance with the Medical Device Directives 90/385/EEC or 93/42/EEC, and which remains valid under the transitional provisions of Article 120(2), (3a) and (3c) of the MDR, can still be submitted to support requirements of Annex I to Directive 2001/83/EC, point 12 of section 3.2, as amended by Article 117 of the MDR.

Article 120(3b) of the MDR provides for a transition period also for class I devices having a declaration of conformity that was drawn up prior to 26 May 2021 under the Directive 93/42/EEC and for which the conformity assessment procedure under the Regulation (EU) 2017/745 requires involvement of a notified body for the first time. If the conditions set out in Article 120(3c) of the MDR are met, the declaration of conformity can be accepted until 31 December 2028 to satisfy requirements of Annex I to Directive 2001/83/EC, point 12 of section 3.2, as amended by Article 117.

2.7. How will the notified body opinion or MAH's GSPRs compliance statement be reflected in the European Public Assessment Report (EPAR)? *Rev. January 2025*

An EPAR provides public information on a medicinal product, including how it was assessed, and reflects the scientific conclusions of the EMA / NCA.

The EPAR will summarise information on the medical device part, relevant to the use with/of the medicinal product, whether a declaration of conformity or, where applicable, an EU certificate or a notified body opinion or MAH's GSPRs compliance statement was submitted as part of the marketing authorisation application for the medicinal product. The applicant/MAH will be requested to identify information in the EPAR that is considered to be commercially confidential and to make a proposal including justifications for deletions/alternative wording. The notified body opinion itself will not be published separately.

2.8. What is the impact of the MDR and Article 117 on marketing authorisation applications of an iDDC on a Mutual Recognition Procedure submitted on or after the 26 May 2021? *Rev. January 2025*

As the Mutual Recognition Procedure (MRP)/Repeat Use Procedure (RUP) is a new application for a marketing authorisation in the concerned Member States, the dossier must comply with the regulatory requirements applicable at the time of the MRP/RUP application. If the requirements have changed since the original National, Decentralised or Mutual Recognition procedure, then the dossier will need to be updated. However, if an application for a iDDC is submitted in the MRP/RUP on or after 26 May 2021 then the General Safety and Performance Requirements (GSPRs) of the MDR and Annex I to Directive 2001/83/EC, point 12 of section 3.2, as amended by Article 117 of the MDR apply and the applicable supporting documentation such as the declaration of conformity, certificate of conformity or notified body opinion or for class I (excluding Is and Im) MAH's compliance statement with the relevant GSPRs of the MDR Annex I must be included in the dossier only in case of a significant change to the design or intended purpose of the device (part), or in case of a new device.

In the absence of device related major changes of the integral DDC authorised in the RMS since the entry into application of the MDR, the previously assessed documentation according to MDD for legacy devices can be accepted.

In case of device changes triggering the need to provide a NB opinion or EU certificate or DoC or MAH's compliance statement, prior to the commencement of the MRP/RUP, a variation application is required to add a declaration of conformity, certificate of conformity or notified body opinion or MAH's GSPRs compliance statement, to formally update the original dossier in the reference Member State and, where applicable, existing concerned Member States.

Prior to the start of the MRP/RUP, the applicant should discuss their plans to provide the required documentation and, as applicable to submit a variation for the existing MAs.

2.9. Do the requirements of MDR Article 117 also apply to an application for medicinal products to be used outside of the European Union (Article 58 or EU-M4all)? *Rev. June 2021*

EMA's Committee for Medicinal Products for Human Use (CHMP) assesses medicines and vaccines under Article 58 of [Regulation \(EC\) No 726/2004](#) to the same rigorous standards as medicines intended for use in the European Union. Therefore, the requirements of Annex I to Directive 2001/83/EC, point 12 of section 3.2, as amended by MDR Article 117 also apply by analogy to Article 58 applications (called EU-M4all).

2.10. Are the requirements for UDI (unique device identifier) applicable to a medicinal product that incorporates, as an integral part, a medical device? *Rev. May 2024*

A iDDC falling under the medicinal products legislation does not have to meet MDR obligations related to UDI. A device part related UDI should therefore not be applied to the package of such a iDDC. Additional information is provided in the [MDCG 2019-2 guidance on application of UDI rules to device-part of products referred to in Article 1\(8\), 1\(9\) and 1\(10\) of Regulation 2017/745](#).

In cases where the device part of a iDDC is CE marked, the product labelling for the iDDC should follow the labelling requirements for medicinal products as outlined in the QRD (working group on Quality Review of Documents) templates. Where an UDI is already directly marked on the device part, it does not need to be removed. The UDI should not appear on the labelling or outer package of the medicinal product.

3. Medicinal products that include a medical device in the secondary packaging of the marketed medicinal product (co-packaged)

3.1. How will the MDR affect the co-packaged medical device? *Rev. May 2024*

Applicants for marketing authorisations of medicinal products where a medical device (e.g. spoons, measuring cups, inhalers, spacers) is provided within the secondary packaging of the marketed medicinal product (i.e. co-packaged) and does not form an integral product with the medicinal product will need to ensure that their co-packaged medical device is CE marked in accordance with the relevant legislation on medical devices to continue placing the product on the market.

- Certain medical devices can benefit from a transitional period provided for in MDR Article 120(3a) to (3e), which allows devices with a valid certificate or declaration of conformity issued under the Directives 93/42/EEC or 90/385/EEC to be placed on the market in accordance with the extended transitional periods until 31 December 2027 or 31 December 2028, depending on the risk class of the device and provided the relevant conditions are met. The medical device manufacturer will need to comply with certain requirements of the MDR from 26 May 2021 (see FAQ's published by the CAMD MDR/IVDR transition subgroup: [FAQs – MDR Transitional provisions; the European Commission services' Q&A on practical aspects related to the implementation of Regulation \(EU\) 2023/607](#) and relevant MDCG guidance available on the Commission [webpage](#)).
- Self-CE marked Class I devices must be in compliance with the MDR by 26 May 2021. If your self-CE marked Class I device is up-classified by the MDR then the transition period provided for in MDR Article 120(3b) is applicable.

3.2. What requirements for medical device labelling are applicable to medical devices "co-packaged" with medicinal products? *Rev. May 2024*

Co-packaged products need to be distinguished from integral drug-device combinations that form a single integral product governed either by Directive 2001/83/EC or Regulation (EC) No 726/2004 or by Regulation (EU) No 2017/745 (see also Q2.1).

In the case of co-packaged products, the medical device must be in conformity with the MDR⁹. This includes requirements regarding the information to be supplied with the device, which are part of the general safety and performance requirements laid down in Annex I of the MDR.

In accordance with Annex I, Chapter III, 23.1 (b) of Regulation (EU) 2017/745, the information required on the label of the medical device (e.g. CE marking, identification of the device, identification of the manufacturer (and, if applicable of the authorised representative), lot/serial number, UDI carrier etc.) should be provided on the device itself or on its own packaging.

The product information annexes (SmPC, labelling and package leaflet) of the medicinal product which is co-packaged with a medical device, should follow the requirements of Directive 2001/83/EC (see [QRD \(Quality Review of Documents\) templates](#)) and should not include any administrative information such as device manufacturer/ authorised representative, CE mark (incl. NB number), device symbols, UDI or references to device vigilance reporting that is provided on or with the medical device.

In accordance with Annex I, Chapter III, Section 23.1 (d) of Regulation (EU) 2017/745, instructions for use (IFU) for class I and class IIa are not required if such devices can be used safely without any such instructions and unless IFU are required by another provision in Section 23. If IFU are required, they should be provided together with the device.

Relevant information for the use of the co-packaged device, especially if necessary for the intended use of the medicinal product with the device should be included in the appropriate sections of the medicinal product package leaflet and SmPC, as applicable (for details please refer to the SmPC guideline & the [QRD \(Quality Review of Documents\) template](#)).

The MAH may itself be the manufacturer of the co-packaged medical device or assume the responsibility of the device manufacturer in accordance with Article 16(1)(a) of the MDR. In this case, the MAH is responsible for compliance with the MDR during the lifecycle of the device and only the contact details of the MAH need to be provided.

Where the MAH is not the manufacturer of the co-packaged medical device, the medical device manufacturer remains responsible for compliance with the MDR during the lifecycle of the device and needs to be identified on the device label and/or IFU.

3.2.1 If co-packaged medical devices class I and class IIa, are supplied without an individual packaging and it is not technically feasible to implement the labelling requirements on the device itself, what alternative solutions could be considered to display the labelling requirements? New May 2024

In accordance with Annex I, Chapter III, 23.1 (b) of Regulation (EU) 2017/745, the information required on the label of the medical device (e.g. identification of the manufacturer, lot/serial number, etc.) should be provided on the device itself or on its own packaging. However, it is understood that co-packaged medical devices, in particular class I and class IIa (such as dosing devices like measuring spoons, measuring cups, or measuring syringes...), can be provided in bulk by the device manufacturer without an individual packaging and will not contain their own packaging or IFU, pursuant to Annex I, section 23.1(d). These devices are by nature small and direct marking of any information on the device itself can be challenging or technically not feasible.

⁹ In accordance with Article 120(3) to (3e) of the MDR, certain devices that are covered by a valid declaration of conformity or certificate issued in accordance with Directive 90/385/EEC or Directive 93/42/EEC may be placed on the market during a transitional period ending on 31 December 2027 or 31 December 2028, depending on the device's risk.

Considering that the product information annexes (which includes the SmPC, labelling and package leaflet) of the medicinal product must not include the required labelling information of the medical device (see Q&A 3.2), the aim of the below proposed solutions is to provide for an acceptable way to include the latter information.

- A. Provide a separate, additional, leaflet within the packaging of the medicinal product, to provide medical device administrative information as per the MDR. This option will result in two separate leaflets included in the secondary packaging, i.e. the package leaflet (PL) of the medicinal product, as well as the leaflet containing the medical device administrative information. It is recommended to include a cross-reference to the other leaflet to avoid that one of the two leaflets is overlooked.

The above option is not preferred where several devices are co-packaged together with the medicinal product. Having several leaflets for different devices in the same package could be confusing for the end-user.

- B. Attach the leaflet containing the medical device administrative information to the package leaflet of the medicinal product and place it within the secondary package of the medicinal product as one single folded component. To implement this option manufacturers shall consider the following:

- the leaflet containing the medical device administrative information should be clearly differentiated from the package leaflet of the medicinal product. This can be achieved by adding the leaflet containing the medical device administrative information as a tear-off section at the end of the printed package leaflet.
- the product information annexes (SmPC, labelling and package leaflet) of the medicinal product, which is co-packaged with a medical device, should follow the requirements of Directive 2001/83/EC and should not include any administrative information of the device as laid down by MDR.
- the purpose of the section containing administrative information of the device should be clearly indicated by using a relevant subheading, e.g. entitled "<Device name> specific information".

- C. Affix a fold out vignette/sticker containing device-specific information directly onto the device itself, or on the packaging of each device, when available.

The following points should be considered when including a fold out vignette/sticker:

- The information on the fold out vignette should be indelible, easily legible, and clearly comprehensible to the intended user or patient; (Article 10(11) MDR).
- The risk of possible loss of information (the sticker can become loose) should be addressed.
- The adhesive should be functional throughout the life cycle of the product (e.g. during shipping, storage in a refrigerator or in a freezer).

3.3. Do I need to submit a declaration of conformity / EU certificate as part of the dossier for a co-packaged medical device? *Rev. May 2024*

It is the responsibility of the applicant/MAH to ensure that the medical devices co-packaged with the medicinal product meet the applicable general safety and performance requirements set out in MDR Annex I and are in compliance with the entirety of the MDR before the combined product is placed on the market. For marketing authorisations of medicinal products co-packaged with a medical device submitted after 26 May 2021, evidence for the device should be provided that relevant requirements

have been met e.g. EU Declaration of Conformity or, where applicable, EU certificate, or other appropriate documentation such as summary information confirming compliance with relevant GSPRs.

For marketing authorisations of medicinal products co-packaged with a medical device submitted or approved prior to 26 May 2021, and where the dossier contains a certificate of conformity or declaration of conformity, it is not necessary to submit a variation to the marketing authorisation to replace the previous evidence of conformity with a new EU certificate or declaration of conformity in compliance with the MDR.

3.4. What actions, if any, do I need to take if my co-packaged device is up-classified and requires to be certified by a notified body for the first time? *Rev. May 2024*

If the medicinal product is co-packaged with a medical device that did not require a notified body assessment under the Medical Device Directive (MDD) but will now require a notified body assessment under the MDR, there is a transition phase up until 31 December 2028 where the device may continue to be placed on the market if it continues to comply with the MDD, provided there are no significant changes in the design and intended purpose of the medical device and other conditions set out in Article 120(3c) of the MDR are met. However, the requirements of the MDR with respect to post-market surveillance, market surveillance, vigilance and registration of economic operators of devices will apply.

4. Lifecycle management

4.1. Will I need to provide a (new or updated) EU declaration of conformity/certificate of conformity issued by a notified body/notified body opinion if there are changes to the device (or device part) after the initial marketing authorisation? *Rev. December 2025*

Contractual agreements between the MAH and the medical device manufacturer should ensure appropriate level of communication and action as regards changes to the device (or device part). As for any other changes, the MAH should determine whether a change to the device – whether integral, co-packaged or, in certain cases, referenced in the product information – may impact the delivery, quality, safety and/or efficacy of the medicine used in combination with the device.

Where a device (or device part) is replaced or a new device is added or where the medical device manufacturer plans to introduce changes that may affect the safety and performance of the device or the conditions prescribed for the intended use of the device, there are three possible situations:

- **For devices covered under a manufacturer's EU declaration of conformity only (no involvement of a notified body):** the device manufacturer is responsible to ensure compliance with the MDR, including changes to the device. The EU declaration of conformity should be updated accordingly, if necessary.
- **For devices covered under an MDR EU certificate issued by a notified body:** if the assessment of changes leads to the issuance of a new/supplemented EU certificate according to the requirements established in the relevant annexes (Annexes IX, X, XI) of the MDR, the EU certificate must be provided as part of an appropriate post-authorisation regulatory procedure.
- **For devices holding a notified body opinion or for class I (excluding Im and Is), a MAH's GSPRs compliance statement:** if the assessment of changes lead to the issuance of a new/updated notified body opinion or MAH's GSPRs compliance statement, the new/updated

notified body opinion or MAH's GSPRs compliance statement must be provided as part of an appropriate post-authorisation regulatory procedure.

Contractual arrangements between the notified body and the device manufacturer must address the notification obligation and the oversight of changes to the device (or device part) in the context of the life-cycle management of an EU certificate.

Article 117 requirement also applies post-authorisation to all marketing authorisations of integral device, irrespective whether already compliant with Annex I to Directive 2001/83/EC (point 12 of section 3.2, as amended by Article 117 MDR) at the time of the initial MAA. For a notified body opinion in the absence of a contractual arrangement between the notified body and the MAH to cover the changes to the device (or device part), it is the responsibility of the MAH to determine when a new or updated NB opinion is needed in support of a change to the device or its intended use. The MAH is responsible for liaising with the competent authorities (i.e. EMA or the relevant NCA) in charge of the iDDC.

4.2. How should I submit changes to the terms of the Marketing Authorisation following changes to the device (or device part)? *Rev. December 2025*

The updated [EC guidelines on the details of the various categories of variations and operation of the procedures](#) (hereafter called 'Variations guidelines'), which applies from 15 January 2026, introduce new variation classifications for submitting changes related to medical devices - whether integral, co-packaged or referenced in the product information devices - under a dedicated section Q.IV. These guidelines define the conditions and documentation to be met for each category of change.

The general principles regarding the lifecycle management of the medicine with regard to device changes when used in combination are summarised below:

- Contractual agreements between the MAH and the medical device manufacturer must ensure appropriate level of communication and action as regards changes to the device part.
- As for any other changes, the MAH should determine whether a change may impact the delivery, quality, safety and/or efficacy of the medicine used in combination with the device.
- If the change affects the registered information or information that is required to be registered in the MA dossier as per the '[QWP-BWP Guideline on medicinal products used with a medical device](#)', a variation application must be submitted in accordance to the Variations guidelines.
- If the change does not impact the registered information but the MAH concludes that there is an impact on the delivery, quality, safety and/or efficacy of the medicine used in combination with the device, a variation application is still required.
- If the change neither affects the registered information nor impacts the delivery, quality, safety, or efficacy, and the information is not required in the MA dossier as per the '[QWP-BWP Guideline on medicinal products used with a medical device](#)', no variation application is needed.

This guidance is provided without prejudice to the fact that assessment of compliance with the GSPRs of the device remains the remit of the notified body or for class I the responsibility of the device manufacturer or of the MAH for integral devices. In case of an EU certificate or device manufacturer's declaration of conformity (DoC), the MAH should follow the MDR requirements for the maintenance of the EU certificate or DoC.

The MAH has the overall responsibility to ensure that the device used in combination with the medicine – in particular for an integral or co-packaged device - is supported by an up-to-date notified body (NB)

opinion or EU certificate or MAH's GSPRs compliance statement. This applies particularly in cases of major changes to the device.

Minor changes to the device (or device part) that do not impact the safety or performance of the device or the intended use of the device but that still require an update of the registered information in the MA dossier should be submitted with the corresponding variation application according to the Variations guidelines. For integral devices, the MAH should assess and provide a justification whether the change has no significant impact on the device to justify the absence of a NB opinion; otherwise, proof of compliance with the MDR should be provided, according to the documentation listed in the Variations guidelines for each type of change.

Of note, changes to medical devices (impacting section 3.2.P.7 or 3.2.R.2 of the MA dossier) should be submitted under the appropriate Q.IV classification, even if the medical device also serves as a container closure system.

For details on the conditions and documentation required for each change category, please refer to the Variations guidelines - Q.IV medical devices.

- **Need for a new or updated notified body opinion**

In order to support identification of situations where a new or revised notified body opinion for the device of an iDDC is anticipated, the following guidance is provided, however this is without prejudice of the specificities of the concerned product case and the fact that assessment of compliance with the general safety and performance requirements (GSPRs) of the device remains the remit of the notified body.

As a guiding approach, a new or revised notified body opinion or for class I (excluding Im and Is), MAH's GSPRs compliance statement for the device of an iDDC is:

- 1) required when a new device is introduced with an extension of MA or a variation;
- 2) expected when major changes are introduced to an existing device, such as:
 - change to its design;
 - addition or replacement of an integral device (or device part)
 - change to its performance characteristics;
 - change to its intended purpose such as a different patient population and/or new user (e.g. home *versus* hospital setting) and/or new usability study, and/or significantly different instructions for use.

which may have a significant impact on the delivery or the quality, safety, or efficacy of the medicinal product.

Of note, in case a change to the device of an iDDC and/or its intended use is (are) not considered to require a new or revised notified body opinion, applicants are expected to provide a justification and risk assessment in Module 3.2.R. Depending on the device change(s) and considering safe and accurate administration and performance, the impact on the Quality Target Product Profile (QTPP), Critical Quality Attributes (CQAs) and control strategy for the iDDC should be considered.

- 3) expected in case of changes to the medicinal product which may impact the performance or safety of a device:

- For example the introduction of a new finished product formulation resulting in different viscosity significantly affecting device performance.
- In case a new pharmaceutical form is associated with the introduction of a new device, depending on whether the medical device forms an integral part with the medicinal product or is co-packaged, the appropriate proof of compliance with the MDR must be provided.

However, in case the new pharmaceutical form is associated with a device that is already authorised for the medicinal product, the MAH should assess whether the new pharmaceutical form may have a significant impact on the safety and performance of the device. In such case a new or updated NB opinion / MAH's GSPRs compliance statement is expected to be provided. In case it is considered not to have significant impact on the safety and performance of the device, a justification should be provided in support to the regulatory procedure to change the MA.

- In case of a new route of administration, it may impact on the safety or performance of the device. In such case a new or updated NB opinion or MAH's compliance statement is likely to be expected in support to the variation to the MA.
- In case of extension of indication e.g. to another condition or new strength to an existing iDDC new and if there is no change to the device or user, the MAH should assess and provide a justification whether the new indication has no significant impact on the safety and performance of the device. In such case a new or updated notified body opinion or MAH's GSPRs compliance statement might not be needed in support to the variation to the MA.

Of note, in the case of medicinal products authorised nationally (i.e. through MRP/DCP or national procedure), a 'line extension' can result in the authorisation of a new marketing authorisation, hence it should be confirmed with the concerned Member States whether this triggers the need for a new notified body opinion *per se*.

In case of doubt regarding the need to provide a new or revised notified body opinion (or DoC or EU Certificate) to support a variation or 'line extension' and where the need for a variation and/or the category of the change is unclear, applicants are strongly recommended to liaise with the national competent authorities for medicinal products well in advance during the pre-submission phase.

To not delay the processing of the variation procedures other than type IA, the (new/updated) EU declaration of conformity /EU certificate issued by a designated notified body / notified body opinion for medical devices should be provided at the time of submission of the variation application. It is highlighted that in any case such documentation is required before CHMP opinion ('line extension' and Type II variations) or Notification (Type IB variations). The notified body opinion should confirm full compliance with the relevant GSPRs and correspond to the intended use of the device of an iDDC as claimed in the 'line extension' or variation.

For type IA variations, this documentation must be provided at time of submission.

5. Consultation procedure for ancillary medicinal substances in medical devices (MDR Art 1(8))

5.1. What type of consultation procedure needs to be carried out for an ancillary medicinal substance that has already been consulted under the medical device Directive 93/42/EEC? Rev. May 2024

According to Article 52(9) MDR, as clarified by [MDCG Guidance 2020-12](#) notified bodies are required to request a consultation with a competent authority for medicinal products as part of the conformity assessment under the MDR for ancillary medicinal substances already consulted under the medical device Directives 93/42/EEC or 90/385/EEC.

It is possible to take the opportunity of an upcoming variation (i.e. change to the ancillary medicinal substance triggering the need for a follow-up consultation opinion) to request an opinion in accordance with the MDR. Please consult the table below to determine whether a full initial consultation or a follow-up (variation) consultation should be submitted.

This guidance is only applicable to the EMA consultation. For nationally competent authorities for medicinal products consultation, please refer to their national guidance or liaise with the NCA.

Table 1. EMA consultation for ancillary medicinal substances under MDR where a consultation already took place under the Directive 93/42/EEC or Directive 90/385/EEC

Procedure type	Timetable	Conditions	Documentation
Type IB variation	First phase 30 days (with possibility to RSI and assessment of responses up to 60 days)	Previous opinion issued by EMA Minor variation (as classified by analogy to Commission Regulation (EC) No 1234/2008) or / and change to NB's assessment of conformity (this includes where the NB is different to MDD consultation)	Full package including description of the manufacturing process and the data relating to the usefulness of incorporation of the substance into the device (according to section 5.2 Annex IX of the MDR)
Type II variation	First phase 60 or 90 days (with possibility to RSI and assessment of responses up to 210 days)	Previous opinion issued by EMA Major variation (as classified by analogy to Commission Regulation (EC) No 1234/2008)	Declaration from manufacturer and NB detailing which elements are changed, if applicable
Full initial consultation	up to 210 days	A different CA has issued the previous opinion	In addition to above: NCA opinion from previous consultation

6. Consultation procedure for companion diagnostics

6.1. What type of consultation procedure needs to be carried out for a companion diagnostic **New May 2024**

A companion diagnostic is an *in vitro* diagnostic test that supports the safe and effective use of a specific medicinal product, by identifying patients that are suitable or unsuitable for treatment.

A companion diagnostic is defined in Article 2(7) of Regulation (EU) 2017/746 as follows: 'companion diagnostic' means a device which is essential for the safe and effective use of a corresponding medicinal product to:

- (a) identify, before and/or during treatment, patients who are most likely to benefit from the corresponding medicinal product; or
- (b) identify, before and/or during treatment, patients likely to be at increased risk of serious adverse reactions as a result of treatment with the corresponding medicinal product.

Annex II of [MDCG 2020-16](#) provides a flowchart to determine whether an *in vitro* diagnostic test fulfils the definition of a companion diagnostic under Regulation (EU) 2017/746.

Regulation (EU) 2017/746 makes companion diagnostics subject to conformity assessment by a notified body, as well as to a consultation of a medicinal products authority regarding the suitability of the device in relation to the medicinal product concerned.

Before it can issue a EU certificate, the notified body must seek a scientific opinion from EMA on the suitability of the companion diagnostic to the medicinal product concerned if:

- the medicinal product falls exclusively within the scope of the centralised procedure for the authorisation of medicines, or
- the medicinal product is already authorised through the centralised procedure, or
- a marketing authorisation application for the medicinal product has been submitted through the centralised procedure.

In other instances, the notified body can seek the opinion either from a national competent authority for medicinal products or from the EMA.

A [procedural guidance](#) is available on the consultation procedure whereby notified bodies seek a scientific opinion from EMA. This is accompanied by a [question-and-answer \(Q&A\) document on practical arrangements](#) and a [frequently asked questions on medicinal products development and assessment involving companion diagnostic \(CDx\)](#).